

FDA Declaration of Compliance

We: **Yokogawa Electric Corporation**
2-9-32 Nakacho, Musashino-shi, Tokyo, 180-8750 Japan

declare that the materials intended to come into contact with food used in the following products comply with the requirements of FDA 21 CFR.

Model code *1	Model name
AXG###	Magnetic Flowmeter
CA####	Magnetic Flowmeter

*1: This FDA declaration of conformity applies to products with the following model, suffix, and option codes:

Model –Suffix and / Option code structure:

AXG### –b c ddd eee f g h i j k -l mm n o /x

Size: ### = 002, 005, 010, 015, 025, 032, 040, 050, 065, 080, 100, 125, 150, 200, 250,
300, 350, 400, and 500

Use: –b = -G, -C, -H or -J

CA#### –b c ddd eee f g h i j k -l mm n o /x

Size: ### = 0015, 0025, 0040, 0050, 0080, 0100, 0150, 0200

Information relevant to the conformity and identification of these Products is provided in bellow.

Parts and materials used in the products

Part	Material Group	Condition of Model – Suffix and Option code	Material
Lining	Plastic	Lining: f = A	PFA
Gasket	Plastic	Lining: f = C Option Code: /BSF, /GF	PTFE
	Rubber	Excluding option code: /GH	EPDM
		Option Code: /GH	Silicon rubber
		Option Code: /GA /GC, /GD,	Fluororubber
		Option Code: /BSC	Chloroprene rubber

Applicable regulations

Material	Regulations
PFA	FDA 21 CFR 177.1550
PTFE	
EPDM	FDA 21 CFR 177.2600
Silicon rubber	
Fluororubber	
Chloroprene rubber	

While the FDA 21 CFR does not explicitly specify metals used in electrodes and grounding devices, or alumina ceramics used in linings, as food-contact materials, we verify the following:

For stainless steel, we confirm that it is manufactured in accordance with ASTM A276/A276M-2015. For titanium and nickel alloys, we confirm that it complies with FDA compliance policy guidelines FDA CPG 545.500.

For alumina ceramics, we use 99.8% purity materials and confirm that they comply with FDA CPG Sec. 545.400 and FDA CPG Sec. 545.450.

We are fully aware of our legal obligations under the above-mentioned FDA Regulation and we will comply with them.

The customer himself is obliged to verify the products with regard to their suitability in the application. This declaration applies to the product and accessories listed in the delivery conditions. This declaration is valid only for standard products manufactured on or before 19 August 2028.

Tokyo, 20 August 2026

Signature:



.....
Soichiro Katsushima
General Manager
Flowmeter Marketing Department
Marketing Division
Sensing Center
Yokogawa Products HQ
Yokogawa Electric Corporation