

Certificate of Compliance

Certificate: 2211690 (LR 96902)

Master Contract: 172608

Project: 2492269

Date Issued: January 20, 2012

Issued to: Yokogawa Electric Corporation
(location M21-5E)
2-9-32 Nakacho
Musashino-shi, Tokyo 180-8750
Japan
Attention: Tooru Akai

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Bill Giesbrecht

Issued by: Bill Giesbrecht, AScT

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Low Concentration Zirconia Oxygen Analyzer, Model OX400 Series, cord connected, rated 100-120 V, 50/60 Hz, 200 VA or 220-240 V, 50/60 Hz, 400 VA or 200-240 V, 50/60 Hz, 400 VA .

Note: The above model is Equipment Class I, Pollution Degree 2, Installation Category II.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 2211690

Master Contract: 172608

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
2492269	January 20, 2012	Update to include alternative power cord set and connector
2456779	September 28, 2011	Update Report 2211690 to include an alternative construction.
2316202	June 14, 2010	Update Report 2211690 to include an alternative construction.
2211690	September 3, 2009	Low Concentration Zirconia Oxygen Analyzer, Model OX400 Series (C/US) per CPC report.